



Certificate of Compliance

Certificate: 2305514

Master Contract: 177541

Project: 2650964

Date Issued: December 16, 2013

Issued to: SCOTT SAFETY

A Div. of Scott Technologies, Inc.
Monroe Corporate Center
4320 Goldmine Rd
Monroe, NC 28110
USA
Attention: Jim Stowe

The products listed below are eligible to bear the CSA Mark shown



Zahra Amini

Issued by: Zahra Amini

PRODUCTS

CLASS 4828 01 -SIGNAL APPLIANCES-Combustible Gas Detection Instruments-For Hazardous Location

Class I, Groups A, B, C and D:

“Protégé” Portable Combustible gas monitor. Intrinsically safe with or without sample pump P/N 096-3259. Temperature Code T4, ambient temperature range -20°C to +50°C. 0-80% LEL.

CLASS 4828 02 -SIGNAL APPLIANCES-Toxic Gas Detection Instruments-For Hazardous Location

Class I, Groups A, B, C and D:

“Protégé” Portable Toxic gas monitor. Intrinsically safe with or without sample pump P/N 096-3259. Temperature Code T4, ambient temperature range -20°C to +50°C.

APPLICABLE REQUIREMENTS

CAN/CSA-C22.2 No. 0-M91 - General Requirements – Canadian Electrical Code, Part II

CSA Std C22.2 No. 152-M1984 - Combustible Gas Detection Instruments

CSA Std C22.2 No. 157-92 - Intrinsically Safe and Non-Incendive Equipment for Use in Hazardous Locations



Supplement to Certificate of Compliance

Certificate: 2305514

Master Contract: 177541

The products listed, including the latest revision described below, are eligible to be marked in accordance with the referenced Certificate.

Product Certification History

Project	Date	Description
2650964	Dec 16, 2013	The Report is updated to add a Toxic version of product and revise version of documents, add product's case color variants, modify LCD holder (potting improvement).
2305514	Mar 23, 2011	Protege Gas Monitor and 096-3259 Sample Pump. Inherently Safe for Class I, Groups A, B, C and D.